

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
 Data File : VN067566.D
 Acq On : 29 Jun 2021 20:33
 Operator : JC/MD
 Sample : M2868-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WASTE-WATER-TRE-1202

Quant Time: Jun 30 03:41:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	178028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	338875	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	324726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	127624	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	136075	55.849	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 111.700%			
35) Dibromofluoromethane	8.024	113	100185	47.684	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 95.360%			
50) Toluene-d8	10.443	98	455460	53.365	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.720%			
62) 4-Bromofluorobenzene	12.734	95	147063	49.330	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 98.660%			
Target Compounds						Qvalue
11) Tert butyl alcohol	5.396	59	25063	68.912	ug/l #	86
16) Acetone	4.293	43	161653	201.147	ug/l	94
18) Methyl Acetate	4.867	43	6518	3.040	ug/l #	77
25) 2-Butanone	7.345	43	25314	20.425	ug/l	98
29) Tetrahydrofuran	7.694	42	79091	95.836	ug/l	90
95) Naphthalene	15.509	128	14953	5.408	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062921\
Data File : VN067566.D
Acq On : 29 Jun 2021 20:33
Operator : JC/MD
Sample : M2868-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WASTE-WATER-TRE-1202

Quant Time: Jun 30 03:41:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

