

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062922\
 Data File : VN073235.D
 Acq On : 29 Jun 2022 11:45
 Operator : JC\MD
 Sample : N3376-01DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 220615003-01-VOADL

Quant Time: Jun 30 06:58:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	333589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	553745	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	458799	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	183612	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	215483	44.746	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.500%		
35) Dibromofluoromethane	7.945	113	164415	44.628	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.260%		
50) Toluene-d8	10.375	98	619970	47.381	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	94.760%		
62) 4-Bromofluorobenzene	12.669	95	209842	41.731	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	83.460%		
Target Compounds						
11) Tert butyl alcohol	5.275	59	237856	203.891	ug/l	99
29) Tetrahydrofuran	7.610	42	40038	13.587	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062922\
Data File : VN073235.D
Acq On : 29 Jun 2022 11:45
Operator : JC\MD
Sample : N3376-01DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
220615003-01-VOADL

Quant Time: Jun 30 06:58:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

