

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062922\
 Data File : VN073247.D
 Acq On : 29 Jun 2022 16:54
 Operator : JC\MD
 Sample : N3504-06 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRE-1618

Quant Time: Jun 29 17:12:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	350140	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	564748	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	513426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	229189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	204373	40.433	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	80.860%		
35) Dibromofluoromethane	7.951	113	177213	47.164	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.320%		
50) Toluene-d8	10.375	98	593169	44.449	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	88.900%		
62) 4-Bromofluorobenzene	12.669	95	249410	48.633	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.260%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.287	59	282957	231.086	ug/l	99
16) Acetone	4.222	43	338503	119.306	ug/l	95
37) Ethyl Acetate	7.339	43	323283m	37.007	ug/l	
40) Benzene	8.392	78	6170	0.346	ug/l	97
52) Toluene	10.445	92	9676	0.878	ug/l	94
67) Ethyl Benzene	11.786	91	4841	0.230	ug/l	98
68) m/p-Xylenes	11.886	106	6472	0.813	ug/l	91
69) o-Xylene	12.216	106	3936	0.485	ug/l	91
84) 1,2,4-Trimethylbenzene	13.310	105	8678	0.539	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062922\
Data File : VN073247.D
Acq On : 29 Jun 2022 16:54
Operator : JC\MD
Sample : N3504-06 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRE-1618

Quant Time: Jun 29 17:12:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

