

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062922\
 Data File : VN073250.D
 Acq On : 29 Jun 2022 18:07
 Operator : JC\MD
 Sample : N3496-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 30 07:00:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	329259	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	537762	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	498080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	247214	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	199483	41.968	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.940%
35) Dibromofluoromethane	7.951	113	171547	47.947	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.900%
50) Toluene-d8	10.380	98	566133	44.552	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.100%
62) 4-Bromofluorobenzene	12.668	95	252429	51.692	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.380%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062922\
Data File : VN073250.D
Acq On : 29 Jun 2022 18:07
Operator : JC\MD
Sample : N3496-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 30 07:00:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

