



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062922\
Data File : VN073229.D
Acq On : 29 Jun 2022 07:49
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Title : SW846 8260
Last Update : Wed Jun 29 01:54:03 2022



AutoFind: Scans 1832, 1833, 1834; Background Corrected with Scan 1823

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	12285	PASS
75	95	30	60	53.8	32952	PASS
95	95	100	100	100.0	61259	PASS
96	95	5	9	6.9	4218	PASS
173	174	0.00	2	0.9	433	PASS
174	95	50	100	76.0	46568	PASS
175	174	5	9	7.9	3668	PASS
176	174	95	101	97.7	45520	PASS
177	176	5	9	6.4	2931	PASS