

Data File : VN049652.D
Acq On : 30 Jun 2018 17:27
Operator : MD\SY
Sample : J3752-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Jul 07 04:08:56 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1307940	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2039448	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1720506	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	619502	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	851452	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
35) Dibromofluoromethane	7.59	113	747833	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
50) Toluene-d8	10.09	98	2777161	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
62) 4-Bromofluorobenzene	12.40	95	785710	38.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.12%	
Target Compounds						
3) Chloromethane	2.06	50	6341	0.344	ug/l	99
16) Acetone	3.83	43	31632	1.166	ug/l	91
20) Methylene Chloride	4.55	84	12586	0.708	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN049652.D
Acq On : 30 Jun 2018 17:27
Operator : MD\SY
Sample : J3752-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Jul 07 04:08:56 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

