

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN063020\
 Data File : VN062237.D
 Acq On : 30 Jun 2020 14:14
 Operator : JC/MD
 Sample : L3151-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-TB-20200626

Quant Time: Jul 01 04:23:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	207192	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	398635	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	381563	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	142262	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	154128	53.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.02%	
35) Dibromofluoromethane	7.55	113	125683	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
50) Toluene-d8	10.06	98	516478	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
62) 4-Bromofluorobenzene	12.38	95	172565	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
Target Compounds						
79) 2-Chlorotoluene	12.65	91	30871	4.054	ug/l	96
82) 4-Chlorotoluene	12.75	91	35096	4.457	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN063020\
Data File : VN062237.D
Acq On : 30 Jun 2020 14:14
Operator : JC/MD
Sample : L3151-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-TB-20200626

Quant Time: Jul 01 04:23:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

