

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN063020\
 Data File : VN062240.D
 Acq On : 30 Jun 2020 15:26
 Operator : JC/MD
 Sample : L3101-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EBS8-W-04

Quant Time: Jul 01 04:30:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	193970	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	377243	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	354370	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	122414	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	149674	54.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.98%	
35) Dibromofluoromethane	7.55	113	120910	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
50) Toluene-d8	10.06	98	481308	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	12.38	95	148401	45.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.80%	
Target Compounds						
79) 2-Chlorotoluene	12.64	91	3202	0.489	ug/l	100
82) 4-Chlorotoluene	12.75	91	3597	0.531	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN063020\
Data File : VN062240.D
Acq On : 30 Jun 2020 15:26
Operator : JC/MD
Sample : L3101-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EBS8-W-04

Quant Time: Jul 01 04:30:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

