

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN063020\
 Data File : VN062242.D
 Acq On : 30 Jun 2020 16:14
 Operator : JC/MD
 Sample : L3127-06 200X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP01-20200625

Quant Time: Jul 01 04:38:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	189482	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	364248	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	335929	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	115614	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	145031	54.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.08%	
35) Dibromofluoromethane	7.55	113	119839	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
50) Toluene-d8	10.06	98	461881	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	12.38	95	142243	45.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.14%	
Target Compounds						
4) Vinyl Chloride	2.16	62	3341	0.991	ug/l	93
27) cis-1,2-Dichloroethene	6.78	96	39003	13.984	ug/l	93
44) Trichloroethene	8.80	130	45108	16.024	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN063020\
Data File : VN062242.D
Acq On : 30 Jun 2020 16:14
Operator : JC/MD
Sample : L3127-06 200X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP01-20200625

Quant Time: Jul 01 04:38:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

