

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN063020\
 Data File : VN062245.D
 Acq On : 30 Jun 2020 17:27
 Operator : JC/MD
 Sample : L3152-03
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-COMP

Quant Time: Jul 01 04:51:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	183546	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	352554	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	321038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	125769	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	139840	54.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.58%	
35) Dibromofluoromethane	7.54	113	110886	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
50) Toluene-d8	10.06	98	437935	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
62) 4-Bromofluorobenzene	12.38	95	135869	44.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.96%	
Target Compounds						
20) Methylene Chloride	4.49	84	3949	1.460	ug/l	94
52) Toluene	10.12	92	102804	15.599	ug/l	98
84) 1,2,4-Trimethylbenzene	13.01	105	9993	1.268	ug/l	94
95) Naphthalene	15.10	128	11107	5.700	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN063020\
Data File : VN062245.D
Acq On : 30 Jun 2020 17:27
Operator : JC/MD
Sample : L3152-03
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-COMP

Quant Time: Jul 01 04:51:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

