

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN063020\
 Data File : VN062248.D
 Acq On : 30 Jun 2020 18:39
 Operator : JC/MD
 Sample : L3170-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1R

Quant Time: Jul 01 05:13:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	197812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	378315	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	362811	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	126394	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	150744	54.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.60%	
35) Dibromofluoromethane	7.55	113	122359	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
50) Toluene-d8	10.06	98	490150	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
62) 4-Bromofluorobenzene	12.38	95	154180	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
Target Compounds						
19) Methyl tert-butyl Ether	5.00	73	36147	4.980	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN063020\
Data File : VN062248.D
Acq On : 30 Jun 2020 18:39
Operator : JC/MD
Sample : L3170-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1R

Quant Time: Jul 01 05:13:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

