

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN063020\
 Data File : VN062254.D
 Acq On : 30 Jun 2020 21:04
 Operator : JC/MD
 Sample : L3170-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2R

Quant Time: Jul 01 05:32:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	187039	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	382139	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	362200	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	128878	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	148301	56.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.00%	
35) Dibromofluoromethane	7.55	113	120406	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
50) Toluene-d8	10.06	98	491505	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
62) 4-Bromofluorobenzene	12.38	95	152148	45.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.90%	
Target Compounds						
11) Tert butyl alcohol	4.73	59	38774	102.887	ug/l #	91
19) Methyl tert-butyl Ether	5.00	73	100996	14.717	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN063020\
Data File : VN062254.D
Acq On : 30 Jun 2020 21:04
Operator : JC/MD
Sample : L3170-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2R

Quant Time: Jul 01 05:32:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

