

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN063020\
 Data File : VN062271.D
 Acq On : 1 Jul 2020 5:07
 Operator : JC/MD
 Sample : L3153-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP2-B

Quant Time: Jul 01 06:13:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	214721	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	438386	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	420252	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	149494	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	167091	55.45	ug/l	0.00
Spiked Amount			Recovery	=	110.90%	
35) Dibromofluoromethane	7.55	113	134289	47.75	ug/l	0.00
Spiked Amount			Recovery	=	95.50%	
50) Toluene-d8	10.06	98	562203	50.01	ug/l	0.00
Spiked Amount			Recovery	=	100.02%	
62) 4-Bromofluorobenzene	12.38	95	180256	47.46	ug/l	0.00
Spiked Amount			Recovery	=	94.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	91075	93.761	ug/l	94
18) Methyl Acetate	4.28	43	6767	2.825	ug/l #	67
20) Methylene Chloride	4.51	84	24917	8.829	ug/l	98
43) Isopropyl Acetate	8.13	43	29407	4.944	ug/l	99
79) 2-Chlorotoluene	12.65	91	8202	1.025	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN063020\
Data File : VN062271.D
Acq On : 1 Jul 2020 5:07
Operator : JC/MD
Sample : L3153-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP2-B

Quant Time: Jul 01 06:13:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

