

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
 Data File : VN067585.D
 Acq On : 30 Jun 2021 20:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jul 01 01:30:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	213618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	404885	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	425433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	220064	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	161282	55.166	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 110.340%	
35) Dibromofluoromethane	8.021	113	121451	48.382	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 96.760%	
50) Toluene-d8	10.443	98	551442	54.077	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 108.160%	
62) 4-Bromofluorobenzene	12.734	95	209362	58.778	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 117.560%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.070	85	99308	44.937	ug/l	98
3) Chloromethane	2.301	50	137448	51.680	ug/l	99
4) Vinyl Chloride	2.440	62	219511	61.387	ug/l	96
5) Bromomethane	2.837	94	162592	61.809	ug/l	98
6) Chloroethane	3.009	64	154114	61.796	ug/l	100
7) Trichlorofluoromethane	3.368	101	404127	62.900	ug/l	99
8) Diethyl Ether	3.819	74	134588	90.406	ug/l	84
9) 1,1,2-Trichlorotrifluo...	4.197	101	93029	43.862	ug/l	97
10) Methyl Iodide	4.411	142	126206	45.652	ug/l	92
11) Tert butyl alcohol	5.390	59	107569	246.489	ug/l	100
12) 1,1-Dichloroethene	4.170	96	91535	44.148	ug/l	84
13) Acrolein	4.036	56	64845	225.757	ug/l	100
14) Allyl chloride	4.832	41	153330	45.238	ug/l #	90
15) Acrylonitrile	5.557	53	282395	251.767	ug/l	99
16) Acetone	4.288	43	213208	221.097	ug/l	93
17) Carbon Disulfide	4.521	76	243945	41.729	ug/l	98
18) Methyl Acetate	4.854	43	120667	46.910	ug/l	93
19) Methyl tert-butyl Ether	5.621	73	399385	50.233	ug/l	97
20) Methylene Chloride	5.093	84	117604	46.465	ug/l	89
21) trans-1,2-Dichloroethene	5.591	96	112549	47.864	ug/l	89
22) Diisopropyl ether	6.498	45	415626	56.470	ug/l	97
23) Vinyl Acetate	6.436	43	1591014	257.854	ug/l #	94
24) 1,1-Dichloroethane	6.385	63	218598	50.895	ug/l	99
25) 2-Butanone	7.340	43	423225	284.591	ug/l	93
26) 2,2-Dichloropropane	7.329	77	186867	46.944	ug/l	93
27) cis-1,2-Dichloroethene	7.327	96	145834	51.268	ug/l	86
28) Bromochloromethane	7.659	49	108192	53.828	ug/l #	80
29) Tetrahydrofuran	7.689	42	287966	290.798	ug/l	88
30) Chloroform	7.820	83	267033	56.515	ug/l	100
31) Cyclohexane	8.094	56	208616	52.478	ug/l	90
32) 1,1,1-Trichloroethane	8.016	97	237593	55.378	ug/l	96
36) 1,1-Dichloropropene	8.222	75	189801	51.926	ug/l	96
37) Ethyl Acetate	7.418	43	176637	52.627	ug/l	97
38) Carbon Tetrachloride	8.206	117	201217	51.389	ug/l	98
39) Methylcyclohexane	9.462	83	211364	47.771	ug/l	91
40) Benzene	8.461	78	581543	52.545	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
 Data File : VN067585.D
 Acq On : 30 Jun 2021 20:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jul 01 01:30:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.640	41	90362	54.006	ug/l #	90
42) 1,2-Dichloroethane	8.531	62	214028	56.027	ug/l	96
43) Isopropyl Acetate	8.566	43	324409	56.579	ug/l #	94
44) Trichloroethene	9.217	130	144342	50.075	ug/l	91
45) 1,2-Dichloropropane	9.494	63	146754	52.966	ug/l	98
46) Dibromomethane	9.580	93	108775	55.140	ug/l	97
47) Bromodichloromethane	9.765	83	215106	55.123	ug/l	98
48) Methyl methacrylate	9.563	41	141141	56.216	ug/l #	86
49) 1,4-Dioxane	9.574	88	51986	989.700	ug/l	88
51) 4-Methyl-2-Pentanone	10.336	43	1037318	299.874	ug/l	93
52) Toluene	10.507	92	432642	59.833	ug/l	100
53) t-1,3-Dichloropropene	10.722	75	242190	52.581	ug/l	99
54) cis-1,3-Dichloropropene	10.191	75	245820	55.557	ug/l #	88
55) 1,1,2-Trichloroethane	10.899	97	163007	57.771	ug/l	97
56) Ethyl methacrylate	10.765	69	260106	60.968	ug/l	86
57) 1,3-Dichloropropane	11.047	76	265634	58.161	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.044	63	278182	268.457	ug/l	92
59) 2-Hexanone	11.089	43	754873	309.098	ug/l	90
60) Dibromochloromethane	11.242	129	182581	53.395	ug/l	100
61) 1,2-Dibromoethane	11.347	107	174646	60.455	ug/l	100
64) Tetrachloroethene	10.980	164	144437	48.454	ug/l	97
65) Chlorobenzene	11.773	112	463552	53.102	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.843	131	174338	55.786	ug/l	100
67) Ethyl Benzene	11.848	91	861826	54.692	ug/l	96
68) m/p-Xylenes	11.956	106	685880	113.058	ug/l	84
69) o-Xylene	12.283	106	335992	55.916	ug/l	85
70) Styrene	12.296	104	561476	52.561	ug/l	92
71) Bromoform	12.460	173	126543	48.465	ug/l #	100
73) Isopropylbenzene	12.581	105	871067	47.070	ug/l	97
74) N-amyl acetate	12.390	43	294962	48.611	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	12.830	83	246325	46.928	ug/l	99
76) 1,2,3-Trichloropropane	12.884	75	304141	68.952	ug/l #	100
77) Bromobenzene	12.865	156	196568	47.875	ug/l	73
78) n-propylbenzene	12.924	91	967145	46.891	ug/l	94
79) 2-Chlorotoluene	13.010	91	593872	47.129	ug/l	89
80) 1,3,5-Trimethylbenzene	13.061	105	729312	48.089	ug/l	93
81) trans-1,4-Dichloro-2-b...	12.626	75	71128	46.361	ug/l	89
82) 4-Chlorotoluene	13.106	91	591816	48.059	ug/l	90
83) tert-Butylbenzene	13.326	119	609458	46.333	ug/l	88
84) 1,2,4-Trimethylbenzene	13.369	105	727226	48.702	ug/l	92
85) sec-Butylbenzene	13.503	105	835379	46.179	ug/l	96
86) p-Isopropyltoluene	13.619	119	694778	47.751	ug/l	93
87) 1,3-Dichlorobenzene	13.619	146	377891	49.708	ug/l	94
88) 1,4-Dichlorobenzene	13.696	146	374996	49.910	ug/l	95
89) n-Butylbenzene	13.943	91	552429	46.037	ug/l	98
90) Hexachloroethane	14.211	117	117943	46.061	ug/l	98
91) 1,2-Dichlorobenzene	13.991	146	363216	50.153	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	14.606	75	45678	49.747	ug/l	67
93) 1,2,4-Trichlorobenzene	15.268	180	152352	47.787	ug/l	100
94) Hexachlorobutadiene	15.373	225	67357	47.938	ug/l	99
95) Naphthalene	15.509	128	524001	48.394	ug/l	100
96) 1,2,3-Trichlorobenzene	15.702	180	148067	47.452	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
Data File : VN067585.D
Acq On : 30 Jun 2021 20:33
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Jul 01 01:30:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.086 min Scan# 2273
Delta R.T. -0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:168 Resp: 213618
Ion Ratio Lower Upper
168 100
99 65.7 62.9 94.3



#2
Dichlorodifluoromethane
Concen: 44.937 ug/l
RT: 2.070 min Scan# 30
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 85 Resp: 99308
Ion Ratio Lower Upper
85 100
87 33.3 16.1 48.2









#3
Chloromethane
Concen: 51.680 ug/l
RT: 2.301 min Scan# 116
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 50 Resp: 137448
Ion Ratio Lower Upper
50 100
52 32.8 25.6 38.4



#4
Vinyl Chloride
Concen: 61.387 ug/l
RT: 2.440 min Scan# 168
Delta R.T. -0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 62 Resp: 219511
Ion Ratio Lower Upper
62 100
64 32.8 24.4 36.6









#5
Bromomethane
Concen: 61.809 ug/l
RT: 2.837 min Scan# 316
Delta R.T. -0.005 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 162592
Ion Ratio Lower Upper
94 100
96 93.6 73.7 110.5



#6
Chloroethane
Concen: 61.796 ug/l
RT: 3.009 min Scan# 380
Delta R.T. -0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 64 Resp: 154114
Ion Ratio Lower Upper
64 100
66 31.1 24.7 37.1









#7
Trichlorofluoromethane
Concen: 62.900 ug/l
RT: 3.368 min Scan# 514
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:101 Resp: 404127
Ion Ratio Lower Upper
101 100
103 64.9 51.2 76.8



#8
Diethyl Ether
Concen: 90.406 ug/l
RT: 3.819 min Scan# 682
Delta R.T. -0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 74 Resp: 134588
Ion Ratio Lower Upper
74 100
45 84.1 50.2 150.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 43.862 ug/l
RT: 4.197 min Scan# 823
Delta R.T. -0.005 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:101 Resp: 93029
Ion Ratio Lower Upper
101 100
85 49.4 39.6 59.4
151 72.2 54.1 81.1



#10
Methyl Iodide
Concen: 45.652 ug/l
RT: 4.411 min Scan# 903
Delta R.T. -0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion:142 Resp: 126206
Ion Ratio Lower Upper
142 100
127 47.0 43.1 64.7
141 14.1 11.4 17.0









#11
Tert butyl alcohol
Concen: 246.489 ug/l
RT: 5.390 min Scan# 1268
Delta R.T. -0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 59 Resp: 107569
Ion Ratio Lower Upper
59 100
57 10.8 8.7 13.1



#12
1,1-Dichloroethene
Concen: 44.148 ug/l
RT: 4.170 min Scan# 813
Delta R.T. -0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 96 Resp: 91535
Ion Ratio Lower Upper
96 100
61 159.5 152.6 229.0
98 63.4 50.9 76.3









#13
Acrolein
Concen: 225.757 ug/l
RT: 4.036 min Scan# 763
Delta R.T. -0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 56 Resp: 64845
Ion Ratio Lower Upper
56 100
55 71.6 57.4 86.0



#14
Allyl chloride
Concen: 45.238 ug/l
RT: 4.832 min Scan# 1060
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 41 Resp: 153330
Ion Ratio Lower Upper
41 100
39 82.8 60.9 91.3
76 38.2 24.3 36.5#









#15
Acrylonitrile
Concen: 251.767 ug/l
RT: 5.557 min Scan# 1330
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 53 Resp: 282395
Ion Ratio Lower Upper
53 100
52 84.4 66.2 99.2
51 37.9 30.2 45.4



#16
Acetone
Concen: 221.097 ug/l
RT: 4.288 min Scan# 857
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 43 Resp: 213208
Ion Ratio Lower Upper
43 100
58 30.3 21.2 31.8









#17
Carbon Disulfide
Concen: 41.729 ug/l
RT: 4.521 min Scan# 944
Delta R.T. -0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 243945
Ion Ratio Lower Upper
76 100
78 9.0 7.7 11.5



#18
Methyl Acetate
Concen: 46.910 ug/l
RT: 4.854 min Scan# 1068
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 43 Resp: 120667
Ion Ratio Lower Upper
43 100
74 24.9 17.4 26.0









#19
Methyl tert-butyl Ether
Concen: 50.233 ug/l
RT: 5.621 min Scan# 1354
Delta R.T. 0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 73 Resp: 399385
Ion Ratio Lower Upper
73 100
57 19.8 17.0 25.6



#20
Methylene Chloride
Concen: 46.465 ug/l
RT: 5.093 min Scan# 1157
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 84 Resp: 117604
Ion Ratio Lower Upper
84 100
49 120.4 110.9 166.3
51 36.7 32.6 49.0
86 61.6 51.6 77.4









#21
trans-1,2-Dichloroethene
Concen: 47.864 ug/l
RT: 5.591 min Scan# 1343
Delta R.T. -0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	96	Resp:	112549
Ion	Ratio	Lower	Upper
96	100		
61	136.4	123.5	185.3
98	62.8	48.6	72.8



#22
Diisopropyl ether
Concen: 56.470 ug/l
RT: 6.498 min Scan# 1681
Delta R.T. -0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion:	45	Resp:	415626
Ion	Ratio	Lower	Upper
45	100		
43	62.6	49.4	74.2
87	31.1	20.9	31.3
59	11.8	9.0	13.4









#23
 Vinyl Acetate
 Concen: 257.854 ug/l
 RT: 6.436 min Scan# 1658
 Delta R.T. 0.000 min
 Lab File: VN067585.D
 Acq: 30 Jun 2021 20:33




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion: 43 Resp: 1591014
 Ion Ratio Lower Upper
 43 100
 86 11.5 7.3 10.9#



#24
 1,1-Dichloroethane
 Concen: 50.895 ug/l
 RT: 6.385 min Scan# 1639
 Delta R.T. 0.000 min
 Lab File: VN067585.D
 Acq: 30 Jun 2021 20:33

Tgt Ion: 63 Resp: 218598
 Ion Ratio Lower Upper
 63 100
 98 5.9 3.0 9.0
 100 4.2 1.9 5.7









#25
2-Butanone
Concen: 284.591 ug/l
RT: 7.340 min Scan# 1995
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 423225
Ion Ratio Lower Upper
43 100
72 28.3 19.7 29.5



#26
2,2-Dichloropropane
Concen: 46.944 ug/l
RT: 7.329 min Scan# 1991
Delta R.T. 0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 77 Resp: 186867
Ion Ratio Lower Upper
77 100
97 22.5 9.8 29.3









#27
 cis-1,2-Dichloroethene
 Concen: 51.268 ug/l
 RT: 7.327 min Scan# 1990
 Delta R.T. -0.000 min
 Lab File: VN067585.D
 Acq: 30 Jun 2021 20:33




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion:	96	Resp:	145834
Ion	Ratio	Lower	Upper
96	100		
61	143.8	0.0	338.8
98	63.9	0.0	129.8



#28
 Bromochloromethane
 Concen: 53.828 ug/l
 RT: 7.659 min Scan# 2114
 Delta R.T. 0.000 min
 Lab File: VN067585.D
 Acq: 30 Jun 2021 20:33

Tgt Ion:	49	Resp:	108192
Ion	Ratio	Lower	Upper
49	100		
129	1.0	0.0	3.8
130	78.3	49.7	74.5#





#29

Tetrahydrofuran

Concen: 290.798 ug/l

RT: 7.689 min Scan# 2125





Delta R.T. -0.000 min

Lab File: VN067585.D

Acq: 30 Jun 2021 20:33

Tgt Ion: 42 Resp: 287966

Ion	Ratio	Lower	Upper
42	100		
72	48.2	32.8	49.2
71	46.3	31.2	46.8



#30

Chloroform

Concen: 56.515 ug/l

RT: 7.820 min Scan# 2174




Delta R.T. 0.000 min


Lab File: VN067585.D

Acq: 30 Jun 2021 20:33

Tgt Ion: 83 Resp: 267033


Ion	Ratio	Lower	Upper
83	100		
85	64.5	51.5	77.3








#31
Cyclohexane
Concen: 52.478 ug/l
RT: 8.094 min Scan# 2276
Delta R.T. -0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33


Instrument :
MSVOA_N
ClientSampled :





Tgt Ion: 56 Resp: 208616
Ion Ratio Lower Upper
56 100
69 33.7 27.3 40.9
84 93.0 64.9 97.3




#32
1,1,1-Trichloroethane
Concen: 55.378 ug/l
RT: 8.016 min Scan# 2247
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33



Tgt Ion: 97 Resp: 237593
Ion Ratio Lower Upper
97 100
99 64.3 50.2 75.2
61 43.9 38.2 57.2









#33
 1,2-Dichloroethane-d4
 Concen: 55.166 ug/l
 RT: 8.437 min Scan# 2404
 Delta R.T. -0.003 min
 Lab File: VN067585.D
 Acq: 30 Jun 2021 20:33




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion: 65 Resp: 161282
 Ion Ratio Lower Upper
 65 100
 67 51.6 0.0 98.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.968 min Scan# 2602
 Delta R.T. 0.000 min
 Lab File: VN067585.D
 Acq: 30 Jun 2021 20:33

Tgt Ion: 114 Resp: 404885
 Ion Ratio Lower Upper
 114 100
 63 20.1 0.0 49.0
 88 15.6 0.0 36.6





#35

Dibromofluoromethane

Concen: 48.382 ug/l

RT: 8.021 min Scan# 2249





Delta R.T. -0.003 min

Lab File: VN067585.D

Acq: 30 Jun 2021 20:33

Tgt Ion:	113	Resp:	121451
Ion Ratio	Lower	Upper	
113	100		
111	103.2	83.8	125.8
192	16.3	13.4	20.0

Instrument :
MSVOA_N
ClientSampleId :



#36

1,1-Dichloropropene

Concen: 51.926 ug/l




RT: 8.222 min Scan# 2324


Delta R.T. 0.000 min

Lab File: VN067585.D

Acq: 30 Jun 2021 20:33

Tgt Ion:	75	Resp:	189801
Ion Ratio	Lower	Upper	
75	100		
110	35.9	15.7	47.0
77	31.0	24.6	37.0





#37

Ethyl Acetate

Concen: 52.627 ug/l

RT: 7.418 min Scan# 2024





Delta R.T. -0.000 min

Lab File: VN067585.D

Acq: 30 Jun 2021 20:33

Tgt Ion:	43	Resp:	176637
Ion	Ratio	Lower	Upper
43	100		
61	14.5	10.6	16.0
70	11.2	8.1	12.1

Instrument :
MSVOA_N
ClientSampleId :



#38

Carbon Tetrachloride

Concen: 51.389 ug/l




RT: 8.206 min Scan# 2318


Delta R.T. 0.000 min

Lab File: VN067585.D

Acq: 30 Jun 2021 20:33

Tgt Ion:	117	Resp:	201217
Ion	Ratio	Lower	Upper
117	100		
119	95.7	75.4	113.2
121	31.3	23.5	35.3





#39

Methylcyclohexane

Concen: 47.771 ug/l

RT: 9.462 min Scan# 2786





Delta R.T. 0.001 min

Lab File: VN067585.D

Acq: 30 Jun 2021 20:33

Tgt Ion: 83 Resp: 211364

Ion	Ratio	Lower	Upper
83	100		
55	72.2	66.6	99.8
98	46.2	35.3	52.9



#40

Benzene

Concen: 52.545 ug/l

RT: 8.461 min Scan# 2413




Delta R.T. 0.000 min


Lab File: VN067585.D

Acq: 30 Jun 2021 20:33

Tgt Ion: 78 Resp: 581543

Ion	Ratio	Lower	Upper
78	100		
77	23.6	18.9	28.3









#41
Methacrylonitrile
Concen: 54.006 ug/l
RT: 7.640 min Scan# 2107
Delta R.T. 0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Instrument :
MSVOA_N
ClientSampled :

Tgt Ion: 41 Resp: 90362




Ion	Ratio	Lower	Upper
41	100		
39	60.0	47.6	71.4
67	87.3	57.7	86.5#
52	36.6	26.2	39.4




#42
1,2-Dichloroethane
Concen: 56.027 ug/l
RT: 8.531 min Scan# 2439
Delta R.T. -0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 62 Resp: 214028

Ion	Ratio	Lower	Upper
62	100		
98	8.5	0.0	14.2









#43
Isopropyl Acetate
Concen: 56.579 ug/l
RT: 8.566 min Scan# 2452
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
43	100		
61	26.6	19.0	28.4
87	14.5	9.5	14.3



#44
Trichloroethene
Concen: 50.075 ug/l
RT: 9.217 min Scan# 2695
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion	Ratio	Lower	Upper
130	100		
95	107.3	0.0	234.4









#45
1,2-Dichloropropane
Concen: 52.966 ug/l
RT: 9.494 min Scan# 2798
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 63 Resp: 146754
Ion Ratio Lower Upper
63 100
65 31.6 24.3 36.5



#46
Dibromomethane
Concen: 55.140 ug/l
RT: 9.580 min Scan# 2830
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 93 Resp: 108775
Ion Ratio Lower Upper
93 100
95 81.8 66.1 99.1
174 81.3 61.0 91.4









#47
Bromodichloromethane
Concen: 55.123 ug/l
RT: 9.765 min Scan# 2899
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 83 Resp: 215106
Ion Ratio Lower Upper
83 100
85 63.9 49.9 74.9
127 8.3 6.1 9.1



#48
Methyl methacrylate
Concen: 56.216 ug/l
RT: 9.563 min Scan# 2824
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 41 Resp: 141141
Ion Ratio Lower Upper
41 100
69 97.3 64.6 96.8#
39 72.7 53.3 79.9





#49

1,4-Dioxane

Concen: 989.700 ug/l

RT: 9.574 min Scan# 2828

Delta R.T. 0.003 min

Lab File: VN067585.D

Acq: 30 Jun 2021 20:33





Tgt Ion: 88 Resp: 51986

Ion Ratio Lower Upper

88 100

43 29.5 28.8 43.2

58 65.5 60.4 90.6



#50

Toluene-d8

Concen: 54.077 ug/l

RT: 10.443 min Scan# 3152

Delta R.T. 0.000 min

Lab File: VN067585.D




Acq: 30 Jun 2021 20:33


Tgt Ion: 98 Resp: 551442

Ion Ratio Lower Upper

98 100

100 63.5 51.7 77.5









#51
4-Methyl-2-Pentanone
Concen: 299.874 ug/l
RT: 10.336 min Scan# 3112
Delta R.T. 0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 1037318
Ion Ratio Lower Upper
43 100
58 39.8 28.6 43.0



#52
Toluene
Concen: 59.833 ug/l
RT: 10.507 min Scan# 3176
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 92 Resp: 432642
Ion Ratio Lower Upper
92 100
91 174.2 139.6 209.4









#53
t-1,3-Dichloropropene
Concen: 52.581 ug/l
RT: 10.722 min Scan# 3256
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 75 Resp: 242190
Ion Ratio Lower Upper
75 100
77 30.9 25.1 37.7



#54
cis-1,3-Dichloropropene
Concen: 55.557 ug/l
RT: 10.191 min Scan# 3058
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 75 Resp: 245820
Ion Ratio Lower Upper
75 100
77 30.8 24.6 37.0
39 50.1 50.7 76.1#









#55
1,1,2-Trichloroethane
Concen: 57.771 ug/l
RT: 10.899 min Scan# 3322
Delta R.T. -0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 97	Resp: 163007
Ion Ratio	Lower Upper
97 100	
83 86.0	72.0 108.0
85 56.3	46.8 70.2
99 60.3	49.8 74.6



#56
Ethyl methacrylate
Concen: 60.968 ug/l
RT: 10.765 min Scan# 3272
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 69	Resp: 260106
Ion Ratio	Lower Upper
69 100	
41 60.7	58.6 88.0
39 39.1	38.3 57.5









#57
1,3-Dichloropropane
Concen: 58.161 ug/l
RT: 11.047 min Scan# 3377
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 265634
Ion Ratio Lower Upper
76 100
78 32.2 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 268.457 ug/l
RT: 10.044 min Scan# 3003
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 63 Resp: 278182
Ion Ratio Lower Upper
63 100
106 28.6 19.8 29.8









#59
2-Hexanone
Concen: 309.098 ug/l
RT: 11.089 min Scan# 3393
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 754873
Ion Ratio Lower Upper
43 100
58 56.1 24.7 74.1



#60
Dibromochloromethane
Concen: 53.395 ug/l
RT: 11.242 min Scan# 3450
Delta R.T. 0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion:129 Resp: 182581
Ion Ratio Lower Upper
129 100
127 77.2 38.6 115.8









#61
1,2-Dibromoethane
Concen: 60.455 ug/l
RT: 11.347 min Scan# 3489
Delta R.T. -0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 174646
Ion Ratio Lower Upper
107 100
109 94.0 75.0 112.4



#62
4-Bromofluorobenzene
Concen: 58.778 ug/l
RT: 12.734 min Scan# 4006
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 95 Resp: 209362
Ion Ratio Lower Upper
95 100
174 74.0 0.0 122.0
176 70.5 0.0 117.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.747 min Scan# 3638
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:117 Resp: 425433
Ion Ratio Lower Upper
117 100
82 56.8 53.1 79.7
119 32.7 25.5 38.3



#64
Tetrachloroethene
Concen: 48.454 ug/l
RT: 10.980 min Scan# 3352
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion:164 Resp: 144437
Ion Ratio Lower Upper
164 100
166 131.0 102.1 153.1
129 106.5 86.9 130.3
131 101.2 83.6 125.4









#65
Chlorobenzene
Concen: 53.102 ug/l
RT: 11.773 min Scan# 3648
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:112 Resp: 463552
Ion Ratio Lower Upper
112 100
114 32.2 24.6 37.0



#66
1,1,1,2-Tetrachloroethane
Concen: 55.786 ug/l
RT: 11.843 min Scan# 3674
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion:131 Resp: 174338
Ion Ratio Lower Upper
131 100
133 95.5 47.5 142.5
119 66.5 33.1 99.3









#67
Ethyl Benzene
Concen: 54.692 ug/l
RT: 11.848 min Scan# 3676
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 861826
Ion Ratio Lower Upper
91 100
106 31.6 23.4 35.0



#68
m/p-Xylenes
Concen: 113.058 ug/l
RT: 11.956 min Scan# 3716
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion:106 Resp: 685880
Ion Ratio Lower Upper
106 100
91 199.4 180.6 270.8









#69
o-Xylene
Concen: 55.916 ug/l
RT: 12.283 min Scan# 3838
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 335992
Ion Ratio Lower Upper
106 100
91 211.5 118.3 354.9



#70
Styrene
Concen: 52.561 ug/l
RT: 12.296 min Scan# 3843
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion:104 Resp: 561476
Ion Ratio Lower Upper
104 100
78 49.5 48.2 72.2
103 55.8 44.5 66.7









#71
Bromoform
Concen: 48.465 ug/l
RT: 12.460 min Scan# 3904
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 126543
Ion Ratio Lower Upper
173 100
175 48.2 23.9 71.8
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.678 min Scan# 4358
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion:152 Resp: 220064
Ion Ratio Lower Upper
152 100
115 60.0 36.1 108.5
150 175.0 0.0 351.8









#73
Isopropylbenzene
Concen: 47.070 ug/l
RT: 12.581 min Scan# 3949
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:105 Resp: 871067
Ion Ratio Lower Upper
105 100
120 26.5 12.5 37.5



#74
N-ethyl acetate
Concen: 48.611 ug/l
RT: 12.390 min Scan# 3878
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 43 Resp: 294962
Ion Ratio Lower Upper
43 100
70 45.8 29.4 44.2#
55 24.3 17.0 25.4
61 25.3 17.6 26.4









#75
1,1,2,2-Tetrachloroethane
Concen: 46.928 ug/l
RT: 12.830 min Scan# 4042
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 246325
Ion Ratio Lower Upper
83 100
131 10.5 4.8 14.4
85 64.0 32.2 96.6



#76
1,2,3-Trichloropropane
Concen: 68.952 ug/l
RT: 12.884 min Scan# 4062
Delta R.T. 0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 75 Resp: 304141
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 47.875 ug/l
RT: 12.865 min Scan# 4055
Delta R.T. 0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 196568
Ion Ratio Lower Upper
156 100
77 205.4 136.2 408.6
158 96.8 48.4 145.3



#78
n-propylbenzene
Concen: 46.891 ug/l
RT: 12.924 min Scan# 4077
Delta R.T. 0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 91 Resp: 967145
Ion Ratio Lower Upper
91 100
120 23.7 10.4 31.4









#79
2-Chlorotoluene
Concen: 47.129 ug/l
RT: 13.010 min Scan# 4109
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 593872
Ion Ratio Lower Upper
91 100
126 34.6 14.5 43.5



#80
1,3,5-Trimethylbenzene
Concen: 48.089 ug/l
RT: 13.061 min Scan# 4128
Delta R.T. -0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 105 Resp: 729312
Ion Ratio Lower Upper
105 100
120 49.0 22.2 66.6









#81
trans-1,4-Dichloro-2-butene
Concen: 46.361 ug/l
RT: 12.626 min Scan# 3966
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion	Ratio	Lower	Upper
75	100		
53	93.6	87.6	131.4
89	47.4	37.8	56.8



#82
4-Chlorotoluene
Concen: 48.059 ug/l
RT: 13.106 min Scan# 4145
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion	Ratio	Lower	Upper
91	100		
126	34.8	14.8	44.4









#83
tert-Butylbenzene
Concen: 46.333 ug/l
RT: 13.326 min Scan# 4227
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:119 Resp: 609458
Ion Ratio Lower Upper
119 100
91 65.9 39.5 118.5
134 25.7 12.0 36.0



#84
1,2,4-Trimethylbenzene
Concen: 48.702 ug/l
RT: 13.369 min Scan# 4243
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion:105 Resp: 727226
Ion Ratio Lower Upper
105 100
120 46.0 20.6 61.8









#85
 sec-Butylbenzene
 Concen: 46.179 ug/l
 RT: 13.503 min Scan# 4293
 Delta R.T. 0.000 min
 Lab File: VN067585.D
 Acq: 30 Jun 2021 20:33




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion:105 Resp: 835379
 Ion Ratio Lower Upper
 105 100
 134 19.8 9.0 27.1



#86
 p-Isopropyltoluene
 Concen: 47.751 ug/l
 RT: 13.619 min Scan# 4336
 Delta R.T. 0.001 min
 Lab File: VN067585.D
 Acq: 30 Jun 2021 20:33

Tgt Ion:119 Resp: 694778
 Ion Ratio Lower Upper
 119 100
 134 26.4 12.4 37.4
 91 24.0 14.8 44.3





#87

1,3-Dichlorobenzene

Concen: 49.708 ug/l

RT: 13.619 min Scan# 4336

Delta R.T. 0.001 min

Lab File: VN067585.D

Acq: 30 Jun 2021 20:33





Tgt Ion:146 Resp: 377891

Ion Ratio Lower Upper

146 100

111 40.8 24.3 72.8

148 62.8 32.0 96.0



#88

1,4-Dichlorobenzene

Concen: 49.910 ug/l

RT: 13.696 min Scan# 4365

Delta R.T. 0.000 min

Lab File: VN067585.D

Acq: 30 Jun 2021 20:33




Tgt Ion:146 Resp: 374996


Ion Ratio Lower Upper

146 100

111 40.2 23.4 70.2

148 64.1 31.8 95.3









#89
n-Butylbenzene
Concen: 46.037 ug/l
RT: 13.943 min Scan# 4457
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 552429
Ion Ratio Lower Upper
91 100
92 51.5 25.7 77.1
134 25.8 11.2 33.6



#90
Hexachloroethane
Concen: 46.061 ug/l
RT: 14.211 min Scan# 4557
Delta R.T. -0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 117 Resp: 117943
Ion Ratio Lower Upper
117 100
201 65.2 33.5 100.5









#91
1,2-Dichlorobenzene
Concen: 50.153 ug/l
RT: 13.991 min Scan# 4475
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:146 Resp: 363216
Ion Ratio Lower Upper
146 100
111 42.5 24.9 74.6
148 63.7 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.747 ug/l
RT: 14.606 min Scan# 4704
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion: 75 Resp: 45678
Ion Ratio Lower Upper
75 100
155 82.4 29.8 89.5
157 104.6 37.6 112.8









#93
1,2,4-Trichlorobenzene
Concen: 47.787 ug/l
RT: 15.268 min Scan# 4951
Delta R.T. 0.003 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:180 Resp: 152352
Ion Ratio Lower Upper
180 100
182 94.7 47.5 142.5
145 34.3 17.2 51.6



#94
Hexachlorobutadiene
Concen: 47.938 ug/l
RT: 15.373 min Scan# 4990
Delta R.T. 0.001 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33


Tgt Ion:225 Resp: 67357
Ion Ratio Lower Upper
225 100
223 63.7 31.4 94.0
227 62.7 31.2 93.6








#95
Naphthalene
Concen: 48.394 ug/l
RT: 15.509 min Scan# 5041
Delta R.T. 0.002 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Instrument :
MSVOA_N
ClientSampled :



Tgt Ion:128 Resp: 524001
Ion Ratio Lower Upper
128 100
127 13.1 10.6 15.8
129 11.0 8.6 12.8



#96
1,2,3-Trichlorobenzene
Concen: 47.452 ug/l
RT: 15.702 min Scan# 5113
Delta R.T. 0.000 min
Lab File: VN067585.D
Acq: 30 Jun 2021 20:33

Tgt Ion:180 Resp: 148067
Ion Ratio Lower Upper
180 100
182 94.7 47.8 143.3
145 36.5 18.2 54.6

