

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
 Data File : VN067571.D
 Acq On : 30 Jun 2021 13:22
 Operator : JC/MD
 Sample : VN0630MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0630MBL01

Quant Time: Jun 30 15:39:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	270949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	504085	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	512299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	200045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	208013	56.096	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.200%
35) Dibromofluoromethane	8.024	113	149029	47.685	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.360%
50) Toluene-d8	10.443	98	709807	55.909	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.820%
62) 4-Bromofluorobenzene	12.734	95	227083	51.207	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
Data File : VN067571.D
Acq On : 30 Jun 2021 13:22
Operator : JC/MD
Sample : VN0630MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0630MBL01

Quant Time: Jun 30 15:39:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

