

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
 Data File : VN067577.D
 Acq On : 30 Jun 2021 17:19
 Operator : JC/MD
 Sample : M2882-01
 Misc : 1.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPILL-COMP-31277-AUD-1077

Quant Time: Jul 01 01:29:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	304359	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	566683	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	605158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	350122	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	228631	54.888	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 109.780%		
35) Dibromofluoromethane	8.018	113	164416	46.797	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 93.600%		
50) Toluene-d8	10.443	98	786102	55.078	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 110.160%		
62) 4-Bromofluorobenzene	12.733	95	322425	64.675	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 129.340%		
Target Compounds						
					Qvalue	
16) Acetone	4.309	43	7527	5.478	ug/l	90
18) Methyl Acetate	4.851	43	7611	2.077	ug/l #	82
39) Methylcyclohexane	9.456	83	15002	2.423	ug/l	89
52) Toluene	10.507	92	943949	93.273	ug/l	98
67) Ethyl Benzene	11.848	91	157230	7.015	ug/l	99
68) m/p-Xylenes	11.953	106	286207	33.166	ug/l	88
69) o-Xylene	12.283	106	134086	15.688	ug/l	90
73) Isopropylbenzene	12.583	105	36832	1.251	ug/l	99
78) n-propylbenzene	12.924	91	160391	4.888	ug/l	99
80) 1,3,5-Trimethylbenzene	13.063	105	276326	11.452	ug/l	97
84) 1,2,4-Trimethylbenzene	13.372	105	890326	37.476	ug/l	95
89) n-Butylbenzene	13.943	91	32662	1.711	ug/l #	78
95) Naphthalene	15.509	128	50890	5.943	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
Data File : VN067577.D
Acq On : 30 Jun 2021 17:19
Operator : JC/MD
Sample : M2882-01
Misc : 1.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPILL-COMP-31277-AUD-1077

Quant Time: Jul 01 01:29:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

