

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
 Data File : VN067579.D
 Acq On : 30 Jun 2021 18:05
 Operator : JC/MD
 Sample : M2832-01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3-WC

Manual Integrations
 APPROVED

MMDadoda
 7/2/2021 10:27:42 AM

Quant Time: Jun 30 18:27:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	305134	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	570890	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	597166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	273785	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	235877	56.483	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 112.960%			
35) Dibromofluoromethane	8.019	113	167595	47.350	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 94.700%			
50) Toluene-d8	10.443	98	805706	56.036	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 112.080%			
62) 4-Bromofluorobenzene	12.733	95	290152	57.772	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 115.540%			
Target Compounds						Qvalue
20) Methylene Chloride	5.058	84	3830m	0.858	ug/l	
95) Naphthalene	15.507	128	4357	3.485	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
Data File : VN067579.D
Acq On : 30 Jun 2021 18:05
Operator : JC/MD
Sample : M2832-01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3-WC

Manual Integrations
APPROVED

MMDadoda
7/2/2021 10:27:42 AM

Quant Time: Jun 30 18:27:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

