

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
 Data File : VN067581.D
 Acq On : 30 Jun 2021 18:55
 Operator : JC/MD
 Sample : M2896-03 40X
 Misc : 8.44g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FS-02

Quant Time: Jul 01 01:29:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	288789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	539811	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	586366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	339144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	220853	55.879	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 111.760%		
35) Dibromofluoromethane	8.024	113	159045	47.521	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 95.040%		
50) Toluene-d8	10.443	98	772740	56.838	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 113.680%		
62) 4-Bromofluorobenzene	12.733	95	329555	69.396	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 138.800%#		
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	7074	5.426	ug/l #	84
18) Methyl Acetate	4.865	43	19384	5.574	ug/l	94
25) 2-Butanone	7.345	43	22396	11.140	ug/l	100
29) Tetrahydrofuran	7.697	42	4764	3.559	ug/l	92
40) Benzene	8.458	78	42959	2.911	ug/l	98
52) Toluene	10.507	92	322410	33.444	ug/l	99
59) 2-Hexanone	11.089	43	48247	14.818	ug/l	81
67) Ethyl Benzene	11.848	91	298653	13.751	ug/l	99
68) m/p-Xylenes	11.953	106	318006	38.032	ug/l	88
69) o-Xylene	12.283	106	161744	19.530	ug/l	86
78) n-propylbenzene	12.921	91	122926	3.867	ug/l	97
80) 1,3,5-Trimethylbenzene	13.061	105	105940	4.533	ug/l	94
84) 1,2,4-Trimethylbenzene	13.372	105	398000	17.295	ug/l	95
86) p-Isopropyltoluene	13.613	119	59593	2.658	ug/l	87
89) n-Butylbenzene	13.943	91	69404	3.753	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
Data File : VN067581.D
Acq On : 30 Jun 2021 18:55
Operator : JC/MD
Sample : M2896-03 40X
Misc : 8.44g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FS-02

Quant Time: Jul 01 01:29:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

