

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
 Data File : VN067570.D
 Acq On : 30 Jun 2021 12:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 30 15:38:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.517	0.524	-1.4	80	0.00
3 P	Chloromethane	0.623	0.631	-1.3	85	0.00
4 C	Vinyl Chloride	0.837	0.915	-9.3#	89	0.00
5 T	Bromomethane	0.616	0.744	-20.8	96	0.00
6 T	Chloroethane	0.584	0.650	-11.3	93	0.00
7 T	Trichlorofluoromethane	1.504	1.735	-15.4	94	0.00
8 T	Diethyl Ether	0.348	0.329	5.5	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.496	0.432	12.9	70	0.00
10 T	Methyl Iodide	0.647	0.535	17.3	67	0.00
11 T	Tert butyl alcohol	0.102	0.094	7.8	75	0.00
12 CM	1,1-Dichloroethene	0.485	0.404	16.7#	69	0.00
13 T	Acrolein	0.071	0.061	14.1	75	0.00
14 T	Allyl chloride	0.793	0.720	9.2	74	0.00
15 T	Acrylonitrile	0.263	0.251	4.6	76	0.00
16 T	Acetone	0.226	0.195	13.7	77	0.00
17 T	Carbon Disulfide	1.368	1.128	17.5	67	0.00
18 T	Methyl Acetate	0.602	0.551	8.5	78	0.00
19 T	Methyl tert-butyl Ether	1.861	1.833	1.5	79	0.00
20 T	Methylene Chloride	0.646	0.522	19.2	72	0.00
21 T	trans-1,2-Dichloroethene	0.550	0.510	7.3	74	0.00
22 T	Diisopropyl ether	1.723	1.840	-6.8	85	0.00
23 T	Vinyl Acetate	1.444	1.565	-8.4	85	0.00
24 P	1,1-Dichloroethane	1.005	0.995	1.0	80	0.00
25 T	2-Butanone	0.348	0.366	-5.2	84	0.00
26 T	2,2-Dichloropropane	0.932	0.933	-0.1	80	0.00
27 T	cis-1,2-Dichloroethene	0.666	0.663	0.5	80	0.00
28 T	Bromochloromethane	0.470	0.475	-1.1	87	0.00
29 T	Tetrahydrofuran	0.232	0.245	-5.6	84	0.00
30 C	Chloroform	1.106	1.170	-5.8#	85	0.00
31 T	Cyclohexane	0.995	0.942	5.3	82	0.00
32 T	1,1,1-Trichloroethane	1.004	1.039	-3.5	80	0.00
33 S	1,2-Dichloroethane-d4	0.684	0.734	-7.3	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.310	0.294	5.2	84	0.00
36 T	1,1-Dichloropropene	0.451	0.454	-0.7	84	0.00
37 T	Ethyl Acetate	0.414	0.417	-0.7	83	0.00
38 T	Carbon Tetrachloride	0.484	0.471	2.7	78	0.00
39 T	Methylcyclohexane	0.546	0.528	3.3	79	0.00
40 TM	Benzene	1.367	1.385	-1.3	83	0.00
41 T	Methacrylonitrile	0.207	0.218	-5.3	88	0.00
42 TM	1,2-Dichloroethane	0.472	0.505	-7.0	90	0.00
43 T	Isopropyl Acetate	0.708	0.743	-4.9	86	0.00
44 TM	Trichloroethene	0.356	0.340	4.5	77	0.00
45 C	1,2-Dichloropropane	0.342	0.343	-0.3#	84	0.00
46 T	Dibromomethane	0.244	0.258	-5.7	86	0.00
47 T	Bromodichloromethane	0.482	0.510	-5.8	85	0.00
48 T	Methyl methacrylate	0.310	0.321	-3.5	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
 Data File : VN067570.D
 Acq On : 30 Jun 2021 12:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 30 15:38:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	82	0.00
50 S	Toluene-d8	1.259	1.333	-5.9	94	0.00
51 T	4-Methyl-2-Pentanone	0.427	0.465	-8.9	89	0.00
52 CM	Toluene	0.893	1.010	-13.1#	91	0.00
53 T	t-1,3-Dichloropropene	0.510	0.583	-14.3	91	0.00
54 T	cis-1,3-Dichloropropene	0.546	0.590	-8.1	86	0.00
55 T	1,1,2-Trichloroethane	0.348	0.387	-11.2	90	0.00
56 T	Ethyl methacrylate	0.527	0.602	-14.2	92	0.00
57 T	1,3-Dichloropropane	0.564	0.625	-10.8	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.104	0.125	-20.2	92	0.00
59 T	2-Hexanone	0.302	0.331	-9.6	90	0.00
60 T	Dibromochloromethane	0.370	0.426	-15.1	89	0.00
61 T	1,2-Dibromoethane	0.357	0.410	-14.8	91	0.00
62 S	4-Bromofluorobenzene	0.440	0.494	-12.3	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.350	0.331	5.4	83	0.00
65 PM	Chlorobenzene	1.026	1.015	1.1	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.367	0.377	-2.7	90	0.00
67 C	Ethyl Benzene	1.852	1.878	-1.4#	91	0.00
68 T	m/p-Xylenes	0.713	0.747	-4.8	93	0.00
69 T	o-Xylene	0.706	0.728	-3.1	92	0.00
70 T	Styrene	1.134	1.214	-7.1	93	0.00
71 P	Bromoform	0.257	0.272	-5.8	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	4.205	3.780	10.1	93	0.00
74 T	N-amyl acetate	1.379	1.254	9.1	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.193	1.052	11.8	93	0.00
76 T	1,2,3-Trichloropropane	1.002	0.850	15.2	83	0.00
77 T	Bromobenzene	0.933	0.849	9.0	93	0.00
78 T	n-propylbenzene	4.686	4.268	8.9	93	0.00
79 T	2-Chlorotoluene	2.863	2.562	10.5	94	0.00
80 T	1,3,5-Trimethylbenzene	3.446	3.193	7.3	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.349	0.318	8.9	91	0.00
82 T	4-Chlorotoluene	2.798	2.589	7.5	95	0.00
83 T	tert-Butylbenzene	2.989	2.687	10.1	92	0.00
84 T	1,2,4-Trimethylbenzene	3.393	3.185	6.1	94	0.00
85 T	sec-Butylbenzene	4.110	3.702	9.9	91	0.00
86 T	p-Isopropyltoluene	3.306	3.136	5.1	93	0.00
87 T	1,3-Dichlorobenzene	1.727	1.655	4.2	95	0.00
88 T	1,4-Dichlorobenzene	1.707	1.655	3.0	98	0.00
89 T	n-Butylbenzene	2.726	2.508	8.0	93	0.00
90 T	Hexachloroethane	0.582	0.507	12.9	85	0.00
91 T	1,2-Dichlorobenzene	1.645	1.593	3.2	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.209	0.192	8.1	92	0.00
93 T	1,2,4-Trichlorobenzene	0.724	0.702	3.0	93	0.00
94 T	Hexachlorobutadiene	0.319	0.318	0.3	91	0.00
95 T	Naphthalene	2.271	2.237	1.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
Data File : VN067570.D
Acq On : 30 Jun 2021 12:03
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 30 15:38:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.709	0.670	5.5	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6