

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
 Data File : VN067572.D
 Acq On : 30 Jun 2021 13:47
 Operator : JC/MD
 Sample : VN0630WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0630WBL01

Quant Time: Jun 30 15:39:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 28 14:31:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	266372	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	502291	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	495823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	197144	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	204898	56.205	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.400%
35) Dibromofluoromethane	8.024	113	147495	47.362	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.720%
50) Toluene-d8	10.443	98	693429	54.814	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.620%
62) 4-Bromofluorobenzene	12.734	95	224905	50.897	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.800%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063021\
Data File : VN067572.D
Acq On : 30 Jun 2021 13:47
Operator : JC/MD
Sample : VN0630WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0630WBL01

Quant Time: Jun 30 15:39:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062821W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 28 14:31:03 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.088 min Scan# 2274
Delta R.T. 0.000 min
Lab File: VN067572.D
Acq: 30 Jun 2021 13:47




Instrument :
MSVOA_N
ClientSampleId :
VN0630WBL01


Tgt Ion:168 Resp: 266372
Ion Ratio Lower Upper
168 100
99 65.6 62.9 94.3



#33
1,2-Dichloroethane-d4
Concen: 56.205 ug/l
RT: 8.440 min Scan# 2405
Delta R.T. -0.000 min
Lab File: VN067572.D
Acq: 30 Jun 2021 13:47

Tgt Ion: 65 Resp: 204898
Ion Ratio Lower Upper
65 100
67 51.2 0.0 98.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.971 min Scan# 2603
Delta R.T. 0.003 min
Lab File: VN067572.D
Acq: 30 Jun 2021 13:47




Instrument :
MSVOA_N
ClientSampleId :
VN0630WBL01


Tgt Ion:114 Resp: 502291
Ion Ratio Lower Upper
114 100
63 21.7 0.0 49.0
88 14.7 0.0 36.6



#35
Dibromofluoromethane
Concen: 47.362 ug/l
RT: 8.024 min Scan# 2250
Delta R.T. 0.000 min
Lab File: VN067572.D
Acq: 30 Jun 2021 13:47

Tgt Ion:113 Resp: 147495
Ion Ratio Lower Upper
113 100
111 104.1 83.8 125.8
192 16.2 13.4 20.0









#50
Toluene-d8
Concen: 54.814 ug/l
RT: 10.443 min Scan# 3152
Delta R.T. 0.000 min
Lab File: VN067572.D
Acq: 30 Jun 2021 13:47




Instrument :
MSVOA_N
ClientSampleId :
VN0630WBL01


Tgt Ion: 98 Resp: 693429
Ion Ratio Lower Upper
98 100
100 62.4 51.7 77.5



#62
4-Bromofluorobenzene
Concen: 50.897 ug/l
RT: 12.734 min Scan# 4006
Delta R.T. 0.001 min
Lab File: VN067572.D
Acq: 30 Jun 2021 13:47

Tgt Ion: 95 Resp: 224905
Ion Ratio Lower Upper
95 100
174 71.4 0.0 122.0
176 69.3 0.0 117.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.749 min Scan# 3639
Delta R.T. 0.003 min
Lab File: VN067572.D
Acq: 30 Jun 2021 13:47

Instrument :
MSVOA_N
ClientSampleId :
VN0630WBL01

Tgt Ion:117 Resp: 495823
Ion Ratio Lower Upper
117 100
82 58.7 53.1 79.7
119 31.6 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.678 min Scan# 4358
Delta R.T. 0.001 min
Lab File: VN067572.D
Acq: 30 Jun 2021 13:47

Tgt Ion:152 Resp: 197144
Ion Ratio Lower Upper
152 100
115 62.1 36.1 108.5
150 156.2 0.0 351.8

