

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063022\
 Data File : VN073261.D
 Acq On : 30 Jun 2022 13:22
 Operator : JC\MD
 Sample : N3554-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-SPC-01-20

Quant Time: Jul 01 08:06:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	282496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	464320	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	416748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	188935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	178308	43.723	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.440%
35) Dibromofluoromethane	7.951	113	149603	48.428	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.860%
50) Toluene-d8	10.381	98	489191	44.586	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.180%
62) 4-Bromofluorobenzene	12.669	95	200138	47.467	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.940%
Target Compounds						
					Qvalue	
16) Acetone	4.252	43	4301	1.879	ug/l #	83
17) Carbon Disulfide	4.463	76	3683	0.427	ug/l #	84
27) cis-1,2-Dichloroethene	7.251	96	4016	0.954	ug/l	88
64) Tetrachloroethene	10.910	164	3437	1.010	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063022\
Data File : VN073261.D
Acq On : 30 Jun 2022 13:22
Operator : JC\MD
Sample : N3554-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-01-20

Quant Time: Jul 01 08:06:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

