

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063022\
 Data File : VN073265.D
 Acq On : 30 Jun 2022 14:59
 Operator : JC\MD
 Sample : N3554-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-SPC-01-50

Quant Time: Jul 01 08:07:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	298759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	503252	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	458486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	216635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	189916	44.034	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.060%
35) Dibromofluoromethane	7.951	113	159732	47.707	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.420%
50) Toluene-d8	10.380	98	530777	44.634	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.260%
62) 4-Bromofluorobenzene	12.669	95	227871	49.863	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.720%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063022\
Data File : VN073265.D
Acq On : 30 Jun 2022 14:59
Operator : JC\MD
Sample : N3554-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-01-50

Quant Time: Jul 01 08:07:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

