

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063022\
 Data File : VN073271.D
 Acq On : 30 Jun 2022 17:26
 Operator : JC\MD
 Sample : N3524-04
 Misc : 1.15g/10mL/100uL/5.0mL/MeOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31546

Quant Time: Jul 01 08:08:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.004	168	293613	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.892	114	485042	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	476037	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	257230	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	177974	41.989	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.980%
35) Dibromofluoromethane	7.945	113	149915	46.456	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.920%
50) Toluene-d8	10.374	98	521754	45.523	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.040%
62) 4-Bromofluorobenzene	12.668	95	253746	57.610	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063022\
Data File : VN073271.D
Acq On : 30 Jun 2022 17:26
Operator : JC\MD
Sample : N3524-04
Misc : 1.15g/10mL/100uL/5.0mL/MeOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31546

Quant Time: Jul 01 08:08:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

