

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063022\
 Data File : VN073274.D
 Acq On : 30 Jun 2022 18:39
 Operator : JC\MD
 Sample : N3554-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-SPC-03-40

Quant Time: Jul 01 08:09:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	323679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	535495	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	497611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	249359	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	196770	42.111	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.220%
35) Dibromofluoromethane	7.957	113	166756	46.806	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.620%
50) Toluene-d8	10.381	98	557959	44.095	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.180%
62) 4-Bromofluorobenzene	12.669	95	254678	52.374	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.740%
Target Compounds						
					Qvalue	
16) Acetone	4.252	43	5223	1.991	ug/l #	82
64) Tetrachloroethene	10.922	164	1830	0.450	ug/l #	74
70) Styrene	12.227	104	3751	0.309	ug/l	92
95) Naphthalene	15.433	128	157441	8.192	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063022\
Data File : VN073274.D
Acq On : 30 Jun 2022 18:39
Operator : JC\MD
Sample : N3554-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-03-40

Quant Time: Jul 01 08:09:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

