

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063022\
 Data File : VN073263.D
 Acq On : 30 Jun 2022 14:11
 Operator : JC\MD
 Sample : N3554-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-SPC-01-30

Quant Time: Jul 01 08:07:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	297540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	498787	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	462676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	226355	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	189330	44.078	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.160%
35) Dibromofluoromethane	7.951	113	157262	47.389	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.780%
50) Toluene-d8	10.380	98	522471	44.329	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.660%
62) 4-Bromofluorobenzene	12.669	95	228915	50.540	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.080%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063022\
Data File : VN073263.D
Acq On : 30 Jun 2022 14:11
Operator : JC\MD
Sample : N3554-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-01-30


Quant Time: Jul 01 08:07:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.016 min Scan# 1042
Delta R.T. -0.000 min
Lab File: VN073263.D
Acq: 30 Jun 2022 14:11

Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-01-30







Tgt Ion:168 Resp: 297540
Ion Ratio Lower Upper
168 100
99 60.1 42.3 63.5



#33
1,2-Dichloroethane-d4
Concen: 44.078 ug/l
RT: 8.369 min Scan# 1102
Delta R.T. -0.000 min
Lab File: VN073263.D
Acq: 30 Jun 2022 14:11

Tgt Ion: 65 Resp: 189330
Ion Ratio Lower Upper
65 100
67 52.2 0.0 104.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.904 min Scan# 1193

Delta R.T. 0.006 min

Lab File: VN073263.D

Acq: 30 Jun 2022 14:11

Instrument :

MSVOA_N

ClientSampleId :

152140-SPC-01-30





Tgt Ion:114 Resp: 498787

Ion Ratio Lower Upper

114 100

63 21.5 0.0 40.0

88 16.1 0.0 32.6



#35

Dibromofluoromethane

Concen: 47.389 ug/l

RT: 7.951 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VN073263.D

Acq: 30 Jun 2022 14:11




Tgt Ion:113 Resp: 157262


Ion Ratio Lower Upper

113 100

111 104.9 81.4 122.0

192 18.5 17.0 25.6









#50
Toluene-d8
Concen: 44.329 ug/l
RT: 10.380 min Scan# 1444
Delta R.T. 0.005 min
Lab File: VN073263.D
Acq: 30 Jun 2022 14:11




Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-01-30


Tgt Ion: 98 Resp: 522471
Ion Ratio Lower Upper
98 100
100 66.5 52.6 78.8



#62
4-Bromofluorobenzene
Concen: 50.540 ug/l
RT: 12.669 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN073263.D
Acq: 30 Jun 2022 14:11

Tgt Ion: 95 Resp: 228915
Ion Ratio Lower Upper
95 100
174 75.3 0.0 167.6
176 72.6 0.0 163.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.680 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN073263.D

Acq: 30 Jun 2022 14:11

Instrument :

MSVOA_N

ClientSampleId :

152140-SPC-01-30





Tgt Ion:117 Resp: 462676

Ion Ratio Lower Upper

117 100

82 56.1 44.2 66.2

119 31.9 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.610 min Scan# 1993

Delta R.T. -0.000 min

Lab File: VN073263.D

Acq: 30 Jun 2022 14:11

Tgt Ion:152 Resp: 226355

Ion Ratio Lower Upper

152 100

115 60.0 28.9 86.7

150 156.6 0.0 356.2

