

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063025\
 Data File : VN087244.D
 Acq On : 30 Jun 2025 15:12
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 01 03:20:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:15:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	116340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	208695	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	187120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	94219	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	75706	51.849	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.700%
35) Dibromofluoromethane	8.177	113	66577	49.822	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.640%
50) Toluene-d8	10.565	98	251553	48.296	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.600%
62) 4-Bromofluorobenzene	12.847	95	83970	45.309	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN063025\
Data File : VN087244.D
Acq On : 30 Jun 2025 15:12
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 01 03:20:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:15:00 2025
Response via : Initial Calibration

