

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070119\
 Data File : VN056565.D
 Acq On : 1 Jul 2019 12:27
 Operator : JC/SP
 Sample : VN0701WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0701WBL01

Quant Time: Jul 02 01:36:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	297802	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	513543	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	495098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	122895	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	182780	55.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.76%	
35) Dibromofluoromethane	7.59	113	159318	49.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.46%	
50) Toluene-d8	10.09	98	643065	51.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.18%	
62) 4-Bromofluorobenzene	12.40	95	187679	45.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070119\
Data File : VN056565.D
Acq On : 1 Jul 2019 12:27
Operator : JC/SP
Sample : VN0701WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0701WBL01

Quant Time: Jul 02 01:36:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

