

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070119\
 Data File : VN056568.D
 Acq On : 1 Jul 2019 13:37
 Operator : JC/SP
 Sample : K3158-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-01-062819

Quant Time: Jul 02 01:42:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	310135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	549792	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	557738	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	156349	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	187900	54.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.34%	
35) Dibromofluoromethane	7.59	113	165258	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
50) Toluene-d8	10.09	98	697954	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
62) 4-Bromofluorobenzene	12.40	95	219608	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
Target Compounds						
16) Acetone	3.82	43	16642	9.664	ug/l	98
20) Methylene Chloride	4.55	84	119499	36.570	ug/l	98
43) Isopropyl Acetate	8.17	43	18748	2.557	ug/l #	91
95) Naphthalene	15.13	128	4755	5.840	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070119\
Data File : VN056568.D
Acq On : 1 Jul 2019 13:37
Operator : JC/SP
Sample : K3158-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-01-062819

Quant Time: Jul 02 01:42:57 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

