

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070119\
 Data File : VN056571.D
 Acq On : 1 Jul 2019 14:48
 Operator : JC/SP
 Sample : K3160-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-01-062819-B

Quant Time: Jul 02 01:54:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	314874	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	547852	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	551176	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	155771	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	189520	54.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.62%	
35) Dibromofluoromethane	7.59	113	163360	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
50) Toluene-d8	10.09	98	691054	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
62) 4-Bromofluorobenzene	12.40	95	217422	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
Target Compounds						
16) Acetone	3.82	43	20550	11.753	ug/l	97
20) Methylene Chloride	4.55	84	65774	19.411	ug/l	96
40) Benzene	8.04	78	238070	16.050	ug/l	99
43) Isopropyl Acetate	8.17	43	16163	2.212	ug/l #	90
52) Toluene	10.16	92	65081	7.177	ug/l	94
67) Ethyl Benzene	11.51	91	25530	1.313	ug/l	100
68) m/p-Xylenes	11.62	106	12807	1.797	ug/l	97
95) Naphthalene	15.13	128	6053	6.005	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070119\
Data File : VN056571.D
Acq On : 1 Jul 2019 14:48
Operator : JC/SP
Sample : K3160-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-01-062819-B

Quant Time: Jul 02 01:54:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

