

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062283.D
 Acq On : 1 Jul 2020 11:21
 Operator : JC/MD
 Sample : VN0701MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0701MBL01

Quant Time: Jul 01 15:55:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	182258	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	363318	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	337649	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	111771	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	148129	57.92	ug/l	0.00
Spiked Amount						
			Recovery	=	115.84%	
35) Dibromofluoromethane	7.54	113	120676	51.77	ug/l	0.00
Spiked Amount						
			Recovery	=	103.54%	
50) Toluene-d8	10.06	98	468686	50.30	ug/l	0.00
Spiked Amount						
			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	12.37	95	138722	44.07	ug/l	0.00
Spiked Amount						
			Recovery	=	88.14%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062283.D
Acq On : 1 Jul 2020 11:21
Operator : JC/MD
Sample : VN0701MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0701MBL01

Quant Time: Jul 01 15:55:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

