

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062290.D
 Acq On : 1 Jul 2020 14:29
 Operator : JC/MD
 Sample : L3129-09ME
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11982ME

Quant Time: Jul 01 15:51:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	191001	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	375355	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	357082	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	141422	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	153147	57.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.28%	
35) Dibromofluoromethane	7.54	113	118541	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
50) Toluene-d8	10.06	98	480790	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
62) 4-Bromofluorobenzene	12.38	95	155663	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
Target Compounds						
16) Acetone	3.79	43	10188	8.191	ug/l #	86
20) Methylene Chloride	4.49	84	2889	0.962	ug/l #	89
84) 1,2,4-Trimethylbenzene	13.01	105	4767	0.538	ug/l	100
95) Naphthalene	15.11	128	5885	4.798	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062290.D
Acq On : 1 Jul 2020 14:29
Operator : JC/MD
Sample : L3129-09ME
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11982ME

Quant Time: Jul 01 15:51:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

