

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062291.D
 Acq On : 1 Jul 2020 14:53
 Operator : JC/MD
 Sample : L3151-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-EB-20200626

Quant Time: Jul 02 07:15:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	196232	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	391200	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	388193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	149171	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	159448	57.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.80%	
35) Dibromofluoromethane	7.55	113	125488	50.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.00%	
50) Toluene-d8	10.06	98	508642	50.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.40%	
62) 4-Bromofluorobenzene	12.38	95	176757	52.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.30%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062291.D
Acq On : 1 Jul 2020 14:53
Operator : JC/MD
Sample : L3151-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-EB-20200626

Quant Time: Jul 02 07:15:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

