

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062296.D
 Acq On : 1 Jul 2020 16:54
 Operator : JC/MD
 Sample : L3170-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jul 02 07:26:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	194105	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	385447	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	368061	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	129321	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	156005	57.27	ug/l	0.00
Spiked Amount			Recovery	=	114.54%	
35) Dibromofluoromethane	7.55	113	122143	49.39	ug/l	0.00
Spiked Amount			Recovery	=	98.78%	
50) Toluene-d8	10.06	98	497309	50.31	ug/l	0.00
Spiked Amount			Recovery	=	100.62%	
62) 4-Bromofluorobenzene	12.38	95	156262	46.79	ug/l	0.00
Spiked Amount			Recovery	=	93.58%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062296.D
Acq On : 1 Jul 2020 16:54
Operator : JC/MD
Sample : L3170-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Jul 02 07:26:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

