

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062300.D
 Acq On : 1 Jul 2020 18:31
 Operator : JC/MD
 Sample : L3192-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-FB-20200630

Quant Time: Jul 02 07:36:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	183805	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	359275	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	341205	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	117733	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	147247	57.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.18%	
35) Dibromofluoromethane	7.55	113	117159	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
50) Toluene-d8	10.06	98	461944	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
62) 4-Bromofluorobenzene	12.38	95	143725	46.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.34%	
Target Compounds						
16) Acetone	3.78	43	9640	7.985	ug/l	99
52) Toluene	10.12	92	40289	5.999	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062300.D
Acq On : 1 Jul 2020 18:31
Operator : JC/MD
Sample : L3192-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-FB-20200630

Quant Time: Jul 02 07:36:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

