

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062305.D
 Acq On : 1 Jul 2020 20:32
 Operator : JC/MD
 Sample : L3127-10 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 15BR-20200625

Quant Time: Jul 02 07:48:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	200457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	385826	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	374113	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	129961	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	154979	55.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.18%	
35) Dibromofluoromethane	7.55	113	126310	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
50) Toluene-d8	10.06	98	502315	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
62) 4-Bromofluorobenzene	12.38	95	159158	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
Target Compounds						
2) Dichlorodifluoromethane	1.83	85	3646	1.609	ug/l	95
4) Vinyl Chloride	2.16	62	59993	16.813	ug/l	98
9) 1,1,2-Trichlorotrifluoroet	3.73	101	4332m	1.881	ug/l	
12) 1,1-Dichloroethene	3.69	96	4964	2.157	ug/l	87
21) trans-1,2-Dichloroethene	4.99	96	5748	2.192	ug/l	97
27) cis-1,2-Dichloroethene	6.78	96	845508	286.557	ug/l	92
44) Trichloroethene	8.80	130	941772	315.848	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062305.D
Acq On : 1 Jul 2020 20:32
Operator : JC/MD
Sample : L3127-10 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
15BR-20200625

Quant Time: Jul 02 07:48:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

