

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062307.D
 Acq On : 1 Jul 2020 21:20
 Operator : JC/MD
 Sample : L3101-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EBS8-W-04

Quant Time: Jul 02 12:29:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	186983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	370385	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	362320	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	132473	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	151355	57.68	ug/l	0.00
Spiked Amount						
			Recovery	=	115.36%	
35) Dibromofluoromethane	7.55	113	121679	51.21	ug/l	0.00
Spiked Amount						
			Recovery	=	102.42%	
50) Toluene-d8	10.06	98	480926	50.63	ug/l	0.00
Spiked Amount						
			Recovery	=	101.26%	
62) 4-Bromofluorobenzene	12.38	95	154681	48.20	ug/l	0.00
Spiked Amount						
			Recovery	=	96.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062307.D
Acq On : 1 Jul 2020 21:20
Operator : JC/MD
Sample : L3101-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EBS8-W-04

Quant Time: Jul 02 12:29:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

