

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062315.D
 Acq On : 2 Jul 2020 2:20
 Operator : JC/MD
 Sample : PB129873ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129873ZHE#01

Quant Time: Jul 02 08:11:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	202936	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	411111	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	415252	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	141029	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	163614	57.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.90%	
35) Dibromofluoromethane	7.55	113	132696	50.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.62%	
50) Toluene-d8	10.06	98	553363	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
62) 4-Bromofluorobenzene	12.38	95	175848	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
Target Compounds						
16) Acetone	3.78	43	12886	10.535	ug/l	100
18) Methyl Acetate	4.28	43	3786m	1.672	ug/l	
20) Methylene Chloride	4.51	84	7627	2.712	ug/l	96
43) Isopropyl Acetate	8.13	43	70637	12.663	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062315.D
Acq On : 2 Jul 2020 2:20
Operator : JC/MD
Sample : PB129873ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129873ZHE#01

Quant Time: Jul 02 08:11:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

