

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062316.D
 Acq On : 2 Jul 2020 2:44
 Operator : JC/MD
 Sample : PB129873ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129873ZHE#02

Quant Time: Jul 02 08:12:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	192250	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	406801	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	408510	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	140639	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	161380	59.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.64%	
35) Dibromofluoromethane	7.55	113	129943	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
50) Toluene-d8	10.06	98	539787	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
62) 4-Bromofluorobenzene	12.38	95	174462	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
Target Compounds						
16) Acetone	3.78	43	12608	11.016	ug/l	94
18) Methyl Acetate	4.29	43	5730	2.672	ug/l #	79
20) Methylene Chloride	4.50	84	7811	2.950	ug/l	94
43) Isopropyl Acetate	8.13	43	70155	12.710	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062316.D
Acq On : 2 Jul 2020 2:44
Operator : JC/MD
Sample : PB129873ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129873ZHE#02

Quant Time: Jul 02 08:12:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

