

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062318.D
 Acq On : 2 Jul 2020 3:32
 Operator : JC/MD
 Sample : PB129873ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129873ZHE#04

Quant Time: Jul 02 08:16:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	196005	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	420341	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	406137	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	135994	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	163510	59.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.90%	
35) Dibromofluoromethane	7.55	113	130981	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
50) Toluene-d8	10.06	98	546497	50.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.40%	
62) 4-Bromofluorobenzene	12.38	95	172487	47.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.72%	
Target Compounds						
16) Acetone	3.78	43	12573	10.685	ug/l	96
18) Methyl Acetate	4.28	43	5651	2.584	ug/l	92
20) Methylene Chloride	4.50	84	7785	2.879	ug/l	92
43) Isopropyl Acetate	8.13	43	71248	12.493	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062318.D
Acq On : 2 Jul 2020 3:32
Operator : JC/MD
Sample : PB129873ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129873ZHE#04

Quant Time: Jul 02 08:16:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

