

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062323.D
 Acq On : 2 Jul 2020 5:33
 Operator : JC/MD
 Sample : PB129873ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129873ZHE#09

Quant Time: Jul 02 08:25:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	178532	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	383959	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	378014	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	126932	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	153542	61.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.56%	
35) Dibromofluoromethane	7.55	113	120160	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
50) Toluene-d8	10.06	98	505131	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
62) 4-Bromofluorobenzene	12.38	95	158860	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
Target Compounds						
16) Acetone	3.78	43	12982	12.662	ug/l	93
18) Methyl Acetate	4.28	43	6770	3.399	ug/l #	85
20) Methylene Chloride	4.50	84	7395	3.012	ug/l	91
43) Isopropyl Acetate	8.13	43	70178	13.471	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062323.D
Acq On : 2 Jul 2020 5:33
Operator : JC/MD
Sample : PB129873ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129873ZHE#09

Quant Time: Jul 02 08:25:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

