

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062324.D
 Acq On : 2 Jul 2020 5:57
 Operator : JC/MD
 Sample : PB129873ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129873ZHE#10

Quant Time: Jul 02 08:27:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	197673	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	421153	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	419853	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	143198	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	166501	60.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.04%	
35) Dibromofluoromethane	7.55	113	131828	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
50) Toluene-d8	10.06	98	560438	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
62) 4-Bromofluorobenzene	12.38	95	175694	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
Target Compounds						
16) Acetone	3.78	43	12512	10.489	ug/l	90
18) Methyl Acetate	4.29	43	7142	3.239	ug/l #	83
20) Methylene Chloride	4.50	84	9787	3.642	ug/l #	96
43) Isopropyl Acetate	8.13	43	71215	12.463	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062324.D
Acq On : 2 Jul 2020 5:57
Operator : JC/MD
Sample : PB129873ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129873ZHE#10

Quant Time: Jul 02 08:27:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

