

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062325.D
 Acq On : 2 Jul 2020 6:21
 Operator : JC/MD
 Sample : L3151-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-SOIL-584-588

Quant Time: Jul 02 08:32:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	196555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	413900	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	411195	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	144916	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	164286	59.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.12%	
35) Dibromofluoromethane	7.55	113	131320	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
50) Toluene-d8	10.06	98	540907	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	12.38	95	176538	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
Target Compounds						
16) Acetone	3.78	43	14272	12.638	ug/l	95
18) Methyl Acetate	4.29	43	6985	3.186	ug/l #	86
20) Methylene Chloride	4.50	84	22145	8.566	ug/l	94
43) Isopropyl Acetate	8.13	43	50843	9.053	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062325.D
Acq On : 2 Jul 2020 6:21
Operator : JC/MD
Sample : L3151-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-SOIL-584-588

Quant Time: Jul 02 08:32:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

