

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\  
 Data File : VN062326.D  
 Acq On : 2 Jul 2020 6:46  
 Operator : JC/MD  
 Sample : L3153-41  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 48 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 SP2-K

Quant Time: Jul 02 08:39:03 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N062420W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 24 15:33:00 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 7.63  | 168  | 190979   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.55  | 114  | 402783   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 11.38 | 117  | 396807   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.32 | 152  | 134531   | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 7.99  | 65   | 158138   | 59.01  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 118.02% |          |
| 35) Dibromofluoromethane    | 7.55  | 113  | 126518   | 48.96  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 97.92%  |          |
| 50) Toluene-d8              | 10.06 | 98   | 531079   | 51.41  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 102.82% |          |
| 62) 4-Bromofluorobenzene    | 12.38 | 95   | 165117   | 47.31  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 94.62%  |          |
| Target Compounds            |       |      |          |        |         |          |
| 16) Acetone                 | 3.78  | 43   | 55524    | 62.972 | ug/l    | 94       |
| 18) Methyl Acetate          | 4.29  | 43   | 7302     | 3.427  | ug/l #  | 66       |
| 20) Methylene Chloride      | 4.51  | 84   | 19010    | 7.543  | ug/l    | 98       |
| 43) Isopropyl Acetate       | 8.13  | 43   | 41154    | 7.530  | ug/l    | 97       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\  
Data File : VN062326.D  
Acq On : 2 Jul 2020 6:46  
Operator : JC/MD  
Sample : L3153-41  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 48 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
SP2-K

Quant Time: Jul 02 08:39:03 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_N\METHODS\82N062420W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jun 24 15:33:00 2020  
Response via : Initial Calibration

