

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062327.D
 Acq On : 2 Jul 2020 7:10
 Operator : JC/MD
 Sample : L3153-44
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP2-L

Quant Time: Jul 02 08:41:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	173902	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	378235	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	359062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	117101	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	151269	61.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.96%	
35) Dibromofluoromethane	7.55	113	120493	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
50) Toluene-d8	10.06	98	486060	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
62) 4-Bromofluorobenzene	12.38	95	147588	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	
Target Compounds						
16) Acetone	3.78	43	78818	100.470	ug/l	97
18) Methyl Acetate	4.28	43	7771	4.006	ug/l #	84
20) Methylene Chloride	4.50	84	20967	9.182	ug/l	97
43) Isopropyl Acetate	8.13	43	41001	7.989	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062327.D
Acq On : 2 Jul 2020 7:10
Operator : JC/MD
Sample : L3153-44
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP2-L

Quant Time: Jul 02 08:41:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

