

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062329.D
 Acq On : 2 Jul 2020 7:58
 Operator : JC/MD
 Sample : L3164-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-SOIL-594-598

Quant Time: Jul 02 08:46:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	200283	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	433066	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	422998	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	142043	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	171303	60.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.90%	
35) Dibromofluoromethane	7.55	113	135860	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
50) Toluene-d8	10.06	98	564853	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
62) 4-Bromofluorobenzene	12.38	95	175620	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
Target Compounds						
16) Acetone	3.78	43	14874	13.020	ug/l	94
18) Methyl Acetate	4.28	43	8192	3.666	ug/l #	50
20) Methylene Chloride	4.51	84	14941	5.598	ug/l #	83
43) Isopropyl Acetate	8.13	43	64169	10.921	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062329.D
Acq On : 2 Jul 2020 7:58
Operator : JC/MD
Sample : L3164-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-SOIL-594-598

Quant Time: Jul 02 08:46:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

