

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070120\
 Data File : VN062330.D
 Acq On : 2 Jul 2020 8:22
 Operator : JC/MD
 Sample : L2814-19
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-02-063020

Quant Time: Jul 02 08:51:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	189572	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	416584	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	414935	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	140986	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	162955	61.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.50%	
35) Dibromofluoromethane	7.55	113	128738	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
50) Toluene-d8	10.06	98	545206	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
62) 4-Bromofluorobenzene	12.38	95	174128	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
Target Compounds						
16) Acetone	3.78	43	14633	13.695	ug/l	96
18) Methyl Acetate	4.28	43	7712	3.647	ug/l #	68
20) Methylene Chloride	4.50	84	18438	7.365	ug/l	94
43) Isopropyl Acetate	8.13	43	64967	11.494	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070120\
Data File : VN062330.D
Acq On : 2 Jul 2020 8:22
Operator : JC/MD
Sample : L2814-19
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-02-063020

Quant Time: Jul 02 08:51:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

