

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070121\
 Data File : VN067597.D
 Acq On : 1 Jul 2021 15:30
 Operator : JC/MD
 Sample : M2861-01 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DSN-002-SEWER-OUTLET-002

Manual Integrations
 APPROVED

MMDadoda
 7/5/2021 2:01:14 PM

Quant Time: Jul 02 06:46:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N070121W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 01 12:09:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.659	128	64060	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.968	114	375465	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.747	117	325090	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.437	65	167719	32.399	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	108.000%
60) 4-Bromofluorobenzene	12.736	95	137031	25.911	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	86.367%
63) Toluene-d8	10.443	98	468346	31.485	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	104.933%
Target Compounds						
15) Acetone	4.296	58	14497m	33.358	ug/l	Qvalue
25) Chloroform	7.820	83	23332	2.920	ug/l	99
41) Bromodichloromethane	9.770	83	2593	0.398	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070121\
Data File : VN067597.D
Acq On : 1 Jul 2021 15:30
Operator : JC/MD
Sample : M2861-01 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DSN-002-SEWER-OUTLET-002

Manual Integrations
APPROVED

MMDadoda
7/5/2021 2:01:14 PM

Quant Time: Jul 02 06:46:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N070121W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 01 12:09:22 2021
Response via : Initial Calibration

