

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070121\
 Data File : VN067598.D
 Acq On : 1 Jul 2021 15:54
 Operator : JC/MD
 Sample : M2861-02 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DSN-002-SEWER-OUTLET-002

Quant Time: Jul 02 06:46:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N070121W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 01 12:09:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.665	128	64100	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	365290	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.749	117	322412	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	168014	32.436	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	108.133%
60) 4-Bromofluorobenzene	12.733	95	132953	25.348	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	84.500%
63) Toluene-d8	10.446	98	458765	31.097	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.667%
Target Compounds						
					Qvalue	
15) Acetone	4.293	58	220806	507.758	ug/l	92
25) Chloroform	7.820	83	521471	65.222	ug/l	99
41) Bromodichloromethane	9.765	83	2028	0.320	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070121\
Data File : VN067598.D
Acq On : 1 Jul 2021 15:54
Operator : JC/MD
Sample : M2861-02 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DSN-002-SEWER-OUTLET-002

Quant Time: Jul 02 06:46:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N070121W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 01 12:09:22 2021
Response via : Initial Calibration

