

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070121\
 Data File : VN067600.D
 Acq On : 1 Jul 2021 16:43
 Operator : JC/MD
 Sample : M2861-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DSN-002-SEWER-OUTLET-002

Manual Integrations
 APPROVED

MMDadoda
 7/5/2021 2:01:21 PM

Quant Time: Jul 02 06:47:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N070121W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 01 12:09:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	56066	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	337858	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.749	117	295676	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.442	65	148584	32.796	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	= 109.333%		
60) 4-Bromofluorobenzene	12.736	95	121209	25.199	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	= 84.000%		
63) Toluene-d8	10.443	98	421784	31.175	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	= 103.933%		
Target Compounds						Qvalue
12) Methyl Acetate	4.862	43	7120m	2.037	ug/l	
15) Acetone	4.293	58	63443	166.797	ug/l	85
25) Chloroform	7.823	83	89893	12.854	ug/l	98
41) Bromodichloromethane	9.767	83	22917	3.914	ug/l #	98
53) Dibromochloromethane	11.240	129	6929	1.508	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070121\
Data File : VN067600.D
Acq On : 1 Jul 2021 16:43
Operator : JC/MD
Sample : M2861-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DSN-002-SEWER-OUTLET-002

Manual Integrations
APPROVED

MMDadoda
7/5/2021 2:01:21 PM

Quant Time: Jul 02 06:47:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N070121W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 01 12:09:22 2021
Response via : Initial Calibration

