



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070121\
 Data File : VN067586.D
 Acq On : 1 Jul 2021 9:07
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N070121W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Thu Jul 01 12:09:22 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3990

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.0	19867	PASS
75	95	30	60	58.5	48475	PASS
95	95	100	100	100.0	82925	PASS
96	95	5	9	7.0	5764	PASS
173	174	0.00	2	0.6	412	PASS
174	95	50	100	77.0	63885	PASS
175	174	5	9	8.4	5347	PASS
176	174	95	101	100.3	64064	PASS
177	176	5	9	6.3	4060	PASS